



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

Part Number:	STSS-509CY		
Lot.No.:	202311200101		
Specification:	Volume: <input checked="" type="checkbox"/> 0.5ml <input type="checkbox"/> 1.5ml <input type="checkbox"/> 2.0ml Sterile: <input type="checkbox"/> Y <input checked="" type="checkbox"/> N Tube color: <input checked="" type="checkbox"/> clear <input type="checkbox"/> amber		
	Cap color: <input type="checkbox"/> clear <input type="checkbox"/> red <input type="checkbox"/> blue <input type="checkbox"/> purple <input checked="" type="checkbox"/> yellow <input type="checkbox"/> green <input type="checkbox"/> amber		
	O-ring: <input checked="" type="checkbox"/> Y <input type="checkbox"/> N		
Product Name:	Screw cap tubes		
Package:	500pcs/bag, 10bags/case		
Manufacturer:	PakGent Bioscience (Suzhou) Co., Ltd.		
Inspection Date:	18th, May, 2026	Report Date	18th, May, 2026
Inspection items and result:			
Cosmetic:	No black spots, no cuts,clear and transparent.With correct color identification		
Package:	Accurate quantity and label, inner package without damage		
Dimension:	OK		
Function test:	60kpa test, 15min no leakage		
Endotoxin:	<0.5EU/ml		
RNase/DNase:	DNase≤4x10 ⁻⁶ Kunitz units/μl; RNase≤10 ⁻⁹ Kunitz units/μl		
PCR Inhibitors:	Not detectable		
Human DNA:	<30pg		
Conclusion: Comply with Specification			

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

Title: QA Manager

Signature: Alvin Cao