



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

Part Number:	CT15R-L		
Lot.No.:	N/A		
Specification:	Volume: <input type="checkbox"/> 0.5ml <input type="checkbox"/> 1.5ml <input type="checkbox"/> 2.0ml <input type="checkbox"/> 5.0ml <input checked="" type="checkbox"/> 15ml <input type="checkbox"/> 50ml		
	Package: <input checked="" type="checkbox"/> Rack <input type="checkbox"/> Bulk		
Product Name:	Centrifuge Tube		
Package:	25pcs/rack, 20racks/case		
Shelf Life:	5 years		
Manufacturer:	PakGent Bioscience (Suzhou) Co., Ltd.		
Inspection Date:	8th, May, 2026	Report Date	8th, May, 2026
Inspection items and result:			
Cosmetic:	No black spots, no cuts, clear.		
Package:	Accurate quantity and label, inner package without damage.		
Dimension:	OK		
Function test:	Centrifuge force 15000xg 10min		
Endotoxin:	<0.03EU/ml		
RNase/DNase:	DNase≤4x10 ⁻⁶ Kunitz units/μl; RNase≤10 ⁻⁹ Kunitz units/μl		
ATP:	2×10 ⁻¹² mg/μl		
PCR Inhibitors:	Not detectable		
Human DNA:	<30pg		
Sterile:	Irradiation, SAL10 ⁻⁶		
Conclusion: Comply with Specification			

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

Title: QA Manager

Signature: Alvin Cao

PakGent Bioscience (Suzhou) Co., Ltd.

Tel: +86 (0) 512 6670 0017

Address: Floor 1-3, Building 33, Alloy Industrial Park, Wuzhong District, Suzhou, China.

PakGent Bioscience (Shanghai) Co., Ltd.

Tel: 400-167-1717

Address: Building C, No 888, Lake west road, Lin gang section, Free Trade Zone, Shanghai, China.

PakGent Bioscience (Changshu) Co., Ltd.

Tel: +86 (0) 512 5298 8330

Address: No.9 Lecheng Road, Bixi Street, Changshu, Jiangsu Province, China.