



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

Part Number:	MCT15		
Lot.No.:	202403160101		
Specification:	Volume: <input type="checkbox"/> 0.5ml <input checked="" type="checkbox"/> 1.5ml <input type="checkbox"/> 2.0ml <input type="checkbox"/> 5.0ml <input type="checkbox"/> 15ml <input type="checkbox"/> 50ml Package: <input type="checkbox"/> Rack <input checked="" type="checkbox"/> Bulk		
Product Name:	Centrifuge Tube		
Package:	500pcs/bag, 10bags/case		
Manufacturer:	PakGent Bioscience (Suzhou) Co., Ltd.		
Inspection Date:	14th, May, 2026	Report Date	14th, May, 2026
Inspection items and result:			
Cosmetic:	No black spots, no cuts, clear.		
Package:	Accurate quantity and label, inner package without damage.		
Dimension:	OK		
Function test:	Centrifuge force 25000xg 10min		
Endotoxin:	<0.5EU/ml		
RNase/DNase:	DNase $\leq 4 \times 10^{-6}$ Kunitz units/ μ l; RNase $\leq 10^{-9}$ Kunitz units/ μ l		
PCR Inhibitors:	Not detectable		
Human DNA:	<30pg		
Conclusion: Comply with Specification			

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

Title: QA Manager

Signature: Alvin Cao