



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

Part Number:	EFG-125P		
Lot.No.:	202403160101		
Specification:	Size: <input checked="" type="checkbox"/> 125ml <input type="checkbox"/> 250ml <input type="checkbox"/> 500ml <input type="checkbox"/> 1000ml		
	Bottom shape: <input checked="" type="checkbox"/> Flat <input type="checkbox"/> Baffled		Cap type: <input checked="" type="checkbox"/> Plug cap <input type="checkbox"/> Vent cap
	Bottle Material: <input checked="" type="checkbox"/> PC <input type="checkbox"/> PETG		Cap Material: <input checked="" type="checkbox"/> HDPE
Product Name:	Erlenmeyer Flask		
Package:	24pcs/case		
Shelf Life:	5 years		
Manufacturer:	PakGent Bioscience (Suzhou) Co., Ltd.		
Inspection Date:	10th, May, 2026	Report Date	10th, May, 2026
Inspection items and result:			
Cosmetic:	No black spots, no cuts, clear and transparent		
Package:	Accurate quantity and label, inner package without damage		
Dimension:	OK		
Function test:	N/A		
Endotoxin:	<0.5EU/ml		
RNase/DNase:	DNase ≤ 4x10 ⁻⁶ Kunitz units/μl; RNase ≤ 10 ⁻⁹ Kunitz units/μl		
ATP:	2x10 ⁻¹² mg/μl		
PCR Inhibitors:	Not detectable		
Human DNA:	<30pg		
Sterile:	Irradiation, SAL 10 ⁻⁶		
Conclusion: Comply with Specification			

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

Title: QA Manager

Signature: Alvin Cao