



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

<b>Part Number:</b>	SPTB50		
<b>Lot.No.:</b>	202308120101		
<b>Specification:</b>	Volume: <input type="checkbox"/> 1ml <input type="checkbox"/> 2ml <input type="checkbox"/> 5ml <input type="checkbox"/> 10ml <input type="checkbox"/> 25ml <input checked="" type="checkbox"/> 50ml		
	Construction: <input type="checkbox"/> one piece construction <input checked="" type="checkbox"/> durable ultrasonically welded 2 pieces		
<b>Product Name:</b>	Serological Pipettes		
<b>Package:</b>	25pcs/bag, 8bags/case		
<b>Shelf Life:</b>	5 years		
<b>Manufacturer:</b>	PakGent Bioscience (Suzhou) Co., Ltd.		
<b>Inspection Date:</b>	17th, May, 2026	<b>Report Date</b>	17th, May, 2026
<b>Inspection items and result:</b>			
<b>Cosmetic:</b>	No black spots, no cuts, clear and transparent. With correct color identification		
<b>Package:</b>	Accurate quantity and label, inner package without damage		
<b>Dimension:</b>	OK		
<b>Function test:</b>	Accuracy $\geq 98\%$		
<b>Endotoxin:</b>	$< 0.5 \text{ EU/ml}$		
<b>RNase/DNase:</b>	DNase $\leq 4 \times 10^{-6}$ Kunitz units/ $\mu\text{l}$ ; RNase $\leq 10^{-9}$ Kunitz units/ $\mu\text{l}$		
<b>ATP:</b>	$2 \times 10^{-12} \text{ mg}/\mu\text{l}$		
<b>PCR Inhibitors:</b>	Not detectable		
<b>Human DNA:</b>	$< 30 \text{ pg}$		
<b>Sterile:</b>	Irradiation, SAL $10^{-6}$		
<b>Conclusion: Comply with Specification</b>			

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

**Title: QA Manager**

**Signature: Alvin Cao**