



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

Part Number:	SPTB02		
Lot.No.:	202307160101		
Specification:	Volume: <input type="checkbox"/> 1ml <input checked="" type="checkbox"/> 2ml <input type="checkbox"/> 5ml <input type="checkbox"/> 10ml <input type="checkbox"/> 25ml <input type="checkbox"/> 50ml		
	Construction: <input checked="" type="checkbox"/> one piece construction <input type="checkbox"/> durable ultrasonically welded 2 piece		
Product Name:	Serological Pipettes		
Package:	50pcs/bag, 20bags/case		
Shelf Life:	5 years		
Manufacturer:	PakGent Bioscience (Suzhou) Co., Ltd.		
Inspection Date:	23rd, May, 2026	Report Date	23rd, May, 2026
Inspection items and result:			
Cosmetic:	No black spots, no cuts, clear and transparent. With correct color identification		
Package:	Accurate quantity and label, inner package without damage		
Dimension:	OK		
Function test:	Accuracy $\geq 98\%$		
Endotoxin:	$< 0.5 \text{ EU/ml}$		
RNase/DNase:	DNase $\leq 4 \times 10^{-6}$ Kunitz units/ μl ; RNase $\leq 10^{-9}$ Kunitz units/ μl		
ATP:	$2 \times 10^{-12} \text{ mg}/\mu\text{l}$		
PCR Inhibitors:	Not detectable		
Human DNA:	$< 30 \text{ pg}$		
Sterile:	Irradiation, SAL 10^{-6}		
Conclusion: Comply with Specification			

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

Title: QA Manager

Signature: Alvin Cao